



Contents lists available at ScienceDirect

MethodsX

journal homepage: www.elsevier.com/locate/mex

Corrigendum

Corrigendum to “VISTO: An open-source device to measure exposure time in psychological experiments”

Andrea De Cesarei^{a,*}, Michele Marzocchi^a, Geoffrey R. Loftus^b^a Department of Psychology, University of Bologna, Italy^b Department of Psychology, University of Washington, United States

ARTICLE INFO

Article history: Available online 22 July 2021

In paper “VISTO: An Open-Source Device to Measure Exposure Time in Psychological Experiments”, a device that measures the onset of experimental visual stimuli is described. This device acquires luminance information through a light sensor BPW42, and then records luminance waveforms or onset times. Throughout the paper however, sensor BPW42 was erroneously described as a photodiode, while it is a phototransistor. Moreover, the wiring scheme in Fig. 1 was incomplete.

In the revised Fig. 1, sensor BPW42 is correctly labeled as a phototransistor, and the correct wiring of all components is reported.

The authors would like to apologize for any inconvenience caused.

DOI of original article: [10.1016/j.mex.2021.101427](https://doi.org/10.1016/j.mex.2021.101427)

* Corresponding author.

E-mail address: andrea.decesarei@unibo.it (A. De Cesarei).

<https://doi.org/10.1016/j.mex.2021.101467>

2215-0161/© 2021 The Author(s). Published by Elsevier B.V. This is an open access article under the CC BY license (<http://creativecommons.org/licenses/by/4.0/>)





Fig. 1. Upper panel, A-C: VISTO main components. Lower panel: diagram of the electronic circuit.