



ALMA MATER STUDIORUM
UNIVERSITÀ DI BOLOGNA

ARCHIVIO ISTITUZIONALE
DELLA RICERCA

Alma Mater Studiorum Università di Bologna Archivio istituzionale della ricerca

Heinz body-related interference with leukocyte and erythrocyte variables obtained by an automated hematology analyzer in cats

This is the final peer-reviewed author's accepted manuscript (postprint) of the following publication:

Published Version:

Heinz body-related interference with leukocyte and erythrocyte variables obtained by an automated hematology analyzer in cats / Dondi F.; Vasylyeva K.; Serafini F.; Guarini M.; Troia R.; Giunti M.; Agnoli C..
- In: JOURNAL OF VETERINARY DIAGNOSTIC INVESTIGATION. - ISSN 1040-6387. - STAMPA. - 31:5(2019), pp. 704-713. [10.1177/1040638719863088]

Availability:

This version is available at: <https://hdl.handle.net/11585/707821> since: 2019-12-07

Published:

DOI: <http://doi.org/10.1177/1040638719863088>

Terms of use:

Some rights reserved. The terms and conditions for the reuse of this version of the manuscript are specified in the publishing policy. For all terms of use and more information see the publisher's website.

This item was downloaded from IRIS Università di Bologna (<https://cris.unibo.it/>).
When citing, please refer to the published version.

(Article begins on next page)

This is the final peer-reviewed accepted manuscript of:

Dondi F, Vasylyeva K, Serafini F, Gruarin M, Troia R, Giunti M, Agnoli C. Heinz body-related interference with leukocyte and erythrocyte variables obtained by an automated hematology analyzer in cats. J Vet Diagn Invest. 2019 Sep;31(5):704-713.

The final published version is available online at doi:[10.1177/1040638719863088](https://doi.org/10.1177/1040638719863088)

Rights / License:

The terms and conditions for the reuse of this version of the manuscript are specified in the publishing policy. For all terms of use and more information see the publisher's website.

This item was downloaded from IRIS Università di Bologna (<https://cris.unibo.it/>)

When citing, please refer to the published version.

1 **Heinz body–related interference with leukocyte and erythrocyte variables obtained by an**
2 **automated hematology analyzer in cats**

3
4 Francesco Dondi, Kateryna Vasylyeva, Federica Serafini, Marta Gruarin, Roberta Troia,
5 Massimo Giunti, Chiara Agnoli¹

6
7 Department of Veterinary Medical Sciences, Alma Mater Studiorum–University of Bologna,
8 Ozzano dell’Emilia, Bologna, Italy.

9
10 ¹Corresponding author: Chiara Agnoli, Department of Veterinary Medical Sciences, Alma Mater
11 Studiorum–University of Bologna, Via Tolara di Sopra 50, 40064, Ozzano dell’Emilia (BO),
12 Italy. chiara.agnoli2@unibo.it

13
14 Running head: Heinz body interference on automated hematology analyzer

15

16 **Abstract.** Heinz bodies (HBs) are known to interfere with automated hematology in cats,
17 particularly with the white blood cell (WBC) count. We evaluated the influence of feline HBs on
18 the complete blood count (CBC) results obtained using a flow cytometry–based analyzer. We
19 retrospectively selected cats with circulating HBs and reviewed the results of their CBCs,
20 including red blood cell (RBC) indices, basophil/lobularity (Baso) WBC count (WBCB),
21 peroxidase (Perox) WBC count (WBCP), and cytograms. Based on the presence or absence of
22 HB-related artifacts in their Baso cytogram, cats were grouped into Baso-HBs and HBs groups,
23 respectively, for comparison. The WBCB and WBCP were compared to manual counts of WBCs
24 carried out on blood smears at 400× (MC-WBC). We included 32 cats in our study: 9 of 32 were
25 in the Baso-HBs group, and 23 of 32 were in the HBs group. Baso-HBs cats had a significantly
26 increased HB percentage ($p < 0.001$), WBCB ($p < 0.001$), difference between WBCB and
27 WBCP ($p < 0.001$), lymphocyte count ($p < 0.001$), mean corpuscular hemoglobin concentration
28 ($p < 0.001$), and difference between calculated and measured erythrocyte hemoglobin
29 concentrations ($p < 0.001$) compared to HBs cats. In Baso-HBs cats, the WBCB was
30 significantly higher than the WBCP ($p = 0.02$); no significant difference was detected between
31 the WBCP and the MC-WBC ($p = 0.88$). Evaluation of automated CBC results raised the
32 suspicion of HB-related interference when using a hematology analyzer in cats; hence, blood
33 smear examination remains essential in routine practice.

34

35 **Key words:** feline; hematology cytograms; peroxidase reaction; spurious leukocytosis.

36

37

Introduction

38 Heinz bodies (HBs) are clumps of oxidized and precipitated hemoglobin on the inner surface of
39 the erythrocyte membrane that can be detected frequently in circulating red blood cells (RBCs)
40 in cats.^{5,10} Oxidative damage could arise from exposure to toxins, such as propylene glycol,^{7,27}
41 onion-containing diets,²¹ drugs (such as acetaminophen, methylene blue, and propofol), and
42 other causes of methemoglobinemia.^{2,16,19} Increased formation of HBs may also occur in cats
43 with spontaneous diseases, such as diabetes mellitus,⁶ hyperthyroidism, and lymphoma.⁵

44 Feline erythrocytes, given their peculiar hemoglobin structure, are particularly prone to
45 hemoglobin oxidative damage and HB formation. Moreover, in cats, the spleen is generally
46 unable to remove RBCs containing HBs.⁸ For these reasons, clinically healthy and non-anemic
47 cats might have up to 5% of circulating RBCs with HBs, and their presence may be >50% when
48 disease is present.⁵ HBs can be seen in blood smears routinely stained with Romanowsky
49 staining; however, smears are best evaluated using a supravital dye, such as brilliant cresyl blue
50 (BCB).¹³ To our knowledge, although different thresholds have been proposed to enumerate HBs
51 in feline blood smears,⁵ there are no specific reference intervals.

52 HBs are known to interfere with automated hematology assays in humans^{9,12} and in
53 cats.^{25,28} The presence of numerous HBs in 2 cats gave an erroneous total white blood cell
54 (WBC) count of $>100 \text{ WBCs} \times 10^9/\text{L}$ when using an automated impedance cell counter.²⁸ An
55 erroneously increased leukocyte count of $69.8 \text{ WBC} \times 10^9/\text{L}$, in a basophil reagent channel of a
56 laser flow cytometry hematology analyzer (Technicon H1; Technicon Instruments, Tarrytown,
57 NY), has been reported in the complete blood count (CBC) of a cat with HBs and hemolytic
58 anemia. This finding was associated with peculiar WBC cytograms.²⁵

82 **Case selection**

83 We reviewed hematology data obtained by an automated hematology assay (ADVIA 2120) and
84 the results of blood smear microscopic examination performed at a university veterinary hospital,
85 between January 2015 and December 2017, in order to identify cats with HBs. Medical records
86 of the cases included were additionally examined in order to retrieve clinical and
87 clinicopathologic data including signalment, etiology of HB formation, and numerical and
88 graphical CBC results.

89 Cats with artifactual changes related to HBs in the Baso scattergram, consisting of a cluster
90 of dots lying at the bottom threshold of this cytogram, as described previously,²⁵ were included
91 in the Baso-HBs group (Fig. 1). Cats with HBs in different percentages on blood smear
92 evaluation and without any artifacts in their cytograms were included in the HBs group. The
93 results of multiple CBCs from the same cat in the Baso-HBs group, when available, were
94 categorized in different arbitrary times after inclusion in the study: T1 (1–3 d), T2 (4–7 d), and
95 T3 (8–14 d).

96 **Blood collection and analysis**

97 Blood samples were obtained by venipuncture using a vacuum system (Vacutest Kima,
98 Arzergrande, Italy), and K₃-EDTA blood samples were processed and analyzed within 2 h after
99 collection. The following automated CBC variables were reviewed: Baso total WBC count
100 (WBCB); Perox total WBC count (WBCP); difference between WBCB and WBCP (Δ WBCB-
101 WBCP); percentage of mononuclear cells (MNs) and polymorphonuclear cells (PMNs);
102 percentage and absolute number of neutrophils, lymphocytes, monocytes, eosinophils, basophils,
103 and large unstained cells (LUCs); RBC count; hematocrit value (Hct); mean corpuscular volume
104 (MCV); red cell volume distribution width (RDW); measured hemoglobin (mHb); calculated Hb

105 (cHb); Hb concentration distribution width (HDW); mean corpuscular Hb (MCH); mean
106 corpuscular Hb concentration (MCHC); corpuscular Hb concentration mean (CHCM); difference
107 between MCHC and CHCM (Δ MCHC-CHCM); reticulocyte percentage (RP) and concentration
108 (RC); and platelet count and mean platelet volume (MPV). The Baso, Perox, and RBC
109 cytograms were also reviewed by 2 of the authors (F Dondi, C Agnoli). Hematology system flags
110 were retrieved from the automated analyzer database; in particular, the percentage of cases with
111 Baso no valley (B-NV; no valid separation between the MN and PMN population in the Baso
112 cytogram), Perox no valley (PX-NV; no valid separation between the noise and lymphocyte
113 population in the Perox cytogram), and a comparison error MCHC-CHCM (CHCMCE) >19 g/L,
114 following the manufacturer's indications for humans (Bayer HealthCare, ADVIA 2120/2120i
115 Operator's guide, 2010), were compared between groups.

116 Blood smears were stained with May Grünwald–Giemsa (MGG; Merck KGaA, Darmstadt,
117 Germany) and were evaluated microscopically. Manual total and differential leukocyte counts
118 were carried out¹¹; in particular, the WBC differential count was determined by counting 100
119 consecutive leukocytes at 1,000 \times ; nucleated red blood cells (NRBC) were also enumerated and
120 recorded as NRBCs/100 WBCs, when present.¹ Erythrocyte and leukocyte morphology, and the
121 estimation of platelet number were evaluated at 1,000 \times . The HB percentage on RBCs was
122 estimated on 1,000 RBCs and reported as absolute values and percentage categories: + if <10%
123 (occasional presence); ++ if 10–50% (moderate presence); +++ if >50% (marked presence), as
124 suggested previously.⁵ All manual counts were carried out in a monolayer area of the smears. In
125 the Baso-HBs group, a WBC manual total estimation (MC-WBC) was determined by
126 multiplying the average number of WBCs counted in 10 fields at 400 \times by $1.6 \times 10^9/L$
127 (Christopher MM. Evaluation of the blood smear. Proc 29th World Small Anim Vet Assoc

128 World Congress; Oct 2004; Rhodes, Greece); The HBs were also evaluated and enumerated as a
129 percentage of HBs on 1,000 RBCs, on a supravital BCB-stained blood smear (Fig. 3; Sarstedt,
130 Nümbrecht, Germany).^{5,13}

131 **Statistical analysis**

132 All data were evaluated with descriptive statistics and reported as median and range (minimum–
133 maximum) or as mean \pm standard deviation, based on their distribution. Normality was checked
134 graphically and using the Shapiro–Wilk test. Differences between groups were evaluated using
135 independent sample *t*-test or Mann–Whitney U test for normally and non-normally distributed
136 values, respectively. In the Baso-HBs group, results at different times were compared using
137 Friedman analysis of variance (ANOVA). Categorical data were analyzed using the Fisher exact
138 test or the chi-squared test. Receiver operating characteristic (ROC) curves were used to find the
139 optimal cutoff values of Δ MCHC-CHCM, Δ WBCB-WBCP, and HB% with the maximal sum of
140 sensitivity and specificity for prediction of the presence of the artifactual changes related to HBs
141 in the Baso scattergram (as reported above). The area under the curve (AUC) was calculated. A *p*
142 ≤ 0.05 was considered significant. Statistical analysis was carried out using MedCalc Statistical
143 Software v.18.2.1 (MedCalc Software, Ostend, Belgium).

144 **Results**

145 Thirty-two cats with HBs were included in our study; 9 of 32 in the Baso-HBs group and 23 of
146 32 in the HBs group. The median age was 93 mo (3–192 mo) in the Baso-HBs group and 113 mo
147 (4–312 mo) in the HBs group. In the Baso-HBs group, 5 of 9 cats were female (3 spayed), and 4
148 of 9 were male (3 neutered); in the HBs group, 9 of 23 were female (6 spayed), and 14 of 23
149 were male (13 neutered). The breeds in the Baso-HBs group included European domestic
150 shorthair (8 of 9) and Maine Coon (1 of 9) cats; those in the HBs group included European

151 domestic shorthair (19 of 23), Birman (2 of 23), Maine Coon (1 of 23), and Persian (1 of 23)
152 cats.

153 In the Baso-HBs group, an association with diseases known to promote HB formation
154 was identified in 6 of 9 cats: diabetic ketoacidosis ($n = 3$), chronic enteropathy/lymphoma ($n =$
155 2), and onion poisoning ($n = 1$); in 3 of 9 cats, the etiology of HB formation remained unknown;
156 intoxication ($n = 2$) and trauma ($n = 1$) were suspected. In the HBs group, HB formation was
157 associated with lymphoma ($n = 4$) and diabetic ketoacidosis ($n = 3$); in 16 of 23, the etiology of
158 the HB remained unknown, and these cats were affected by acute or chronic enteropathy ($n = 5$),
159 chronic kidney disease ($n = 3$), feline leukemia virus or feline immunodeficiency virus infection
160 ($n = 2$), diaphragmatic hernia ($n = 1$), Cushing syndrome ($n = 1$), non-regenerative anemia ($n =$
161 1), allergic dermatitis ($n = 1$), neurologic disease ($n = 1$), and trauma ($n = 1$).

162 Complete hematology and ROC curve analysis results are reported in Tables 1–3. The
163 number and percentage of HBs on RBCs were significantly increased in the Baso-HBs group as
164 compared to the HBs group (Table 1). In particular, a marked presence of HBs was detected
165 more frequently in the Baso-HBs group than in the HBs group (Table 1). The Baso-HBs group
166 had significantly increased MCH, MCHC, Δ MCHC-CHCM, RP, and RC when compared to the
167 HBs group (Table 1). The MCHC values were significantly higher as compared with the CHCM
168 in both the Baso-HBs and HBs groups ($p < 0.001$ and $p = 0.01$, respectively; Fig. 4).

169 The cats in the Baso-HBs group had significantly increased WBCB, Δ WBCB-WBCP,
170 PMNs, and absolute and percentage lymphocyte counts, and significantly decreased MNs and
171 neutrophil percentages when compared to the HBs group (Table 2). In the Baso-HBs group, the
172 WBCB was significantly greater than the WBCP ($p = 0.02$), whereas the WBCB and WBCP

Commentato [HF1]: Use of a slash is discouraged in JVDI. Should this be "or", "and/or", or "and"?

173 were not different in the HBs group ($p = 0.88$; Fig. 5). No significant difference was detected
174 between WBCP and MC-WBC within the Baso-HBs group ($p = 0.40$).

175 In the Baso-HBs cats, automated percentage and absolute lymphocyte counts were higher
176 than the manual lymphocyte count ($p < 0.001$ and $p < 0.001$, respectively), and the automated
177 neutrophil percentage count was lower than the manual neutrophil percentage ($p = 0.002$). A
178 similar difference was not noticed for the absolute manual and automated neutrophil counts.

179 All cats included in the Baso-HBs group presented an artifact in the Perox cytogram that
180 was characterized by numerous dots located in the bottom left-hand corner of the graph (Fig. 2).
181 This artifact was not clearly detected in the HBs group. The evaluation of the RBC scattergrams
182 revealed that the Baso-HBs cats had an RBC population right-shifted toward hyperchromic, and
183 occasionally microcytic, erythrocytes (Fig. 6). Similar findings were not detectable in the HBs
184 group. The Baso-HBs cats had a significantly higher frequency of PX-NV, B-NV, and CHCMCE
185 flags as compared to the HBs cats (Tables 1, 2).

186 The hematologic data of the Baso-HBs cats at all of the different times were available in
187 4 of 9 cats (Table 4). Over time, in the Baso-HBs group, there was a significant and progressive
188 decrease in HB percentage. Concurrently, Δ WBCB-WBCP, MCHC, CHCM, and Δ MCHC-
189 CHCM decreased significantly between T0 and T2, and T0 and T3, and between T2 and T3. In 3
190 of 4 cats in the Baso-HBs group, there was complete disappearance of the Baso-channel-related
191 artifactual changes at T3.

192 Discussion

193 We evaluated the potential HB interference with hematology results in cats obtained by an
194 automated analyzer by comparison of numerical and graphical data. During the study period, we
195 evaluated cases in which a marked presence of HBs produced relevant artifacts in the CBC

196 results. Underlying diseases previously related to HB formation were identified by clinicians in
197 the majority (6 of 9) of the Baso-HBs cats; no clear etiology promoting HB formation was
198 identified in the remaining subjects. In particular, to our knowledge, traumatic injuries have
199 never before been associated with HB formation in cats. Although the identified underlying
200 diseases were similar in HBs and Baso-HBs cats, the number of HBs was significantly increased
201 in the latter. In our study, in particular, the diseases more frequently associated with HB presence
202 were diabetic ketoacidosis and lymphoma. A previous study reported a significant correlation
203 between the number of HBs and the plasma concentration of β -hydroxy-butyrate.⁶ Similar
204 findings, however, were not confirmed in our study (data not shown). Different disease severity
205 associated with a more compromised antioxidant status can only be hypothesized to explain the
206 difference detected in our study groups.

207 Our results indicated that an elevated number of HBs in feline erythrocytes significantly
208 affected some hematologic results of the automated analyzer used in our laboratory. In particular,
209 the presence of elevated percentages of HBs (>36% of RBCs) is highly associated with WBCB
210 artifacts and leads to an erroneously increased total WBCB count, which must be verified by
211 blood smear examination.

212 In the hematology system used in our study, the Baso and Perox assays were the default
213 methods for routine total and differential WBC counts, respectively. In humans, as in veterinary
214 species, however, both the Perox and the Baso methods are able to provide total WBC counts
215 (WBCB and WBCP) that must coincide or agree within specific limits.¹⁷ In humans, a
216 discrepancy between the results of the 2 methods, calculated as an increase in the Δ WBCB-
217 WBCP, has been reported in the presence of NRBCs, malaria parasites, or incomplete RBC lysis
218 (e.g., the presence of HBs).²⁹ To our knowledge, the causes of increased Δ WBCB-WBCP are not

219 well known in veterinary medicine. In all of the cats included in our study, an elevated
220 concentration of NRBCs was ruled out by blood smear evaluation. The artifactual leukocytosis
221 detected by the Baso channel and the significantly higher Δ WBCB-WBCP observed in the Baso-
222 HBs cats could be reasonably related to the elevated number of HBs. In particular, WBCB count
223 reached 459 and 549×10^9 WBC/L in 2 cases, further increasing the difference between the
224 study groups. Feline HBs are predominantly distinct, large, round structures and have a high
225 density that frequently leads to incomplete RBC lysis^{20,26}; hence, interference with the Baso
226 method could also have been expected in the populations of our study, as described previously.²⁵
227 Moreover, the events represented as a cluster of small particles overlapping the bottom threshold
228 of the graph could have incremented the WBCB incorrectly and prevented a correct automated
229 separation between MN and PMN cells, flagged by B-NV. Visual evaluation of the Baso
230 cytograms carried out by the operator would have allowed identification of the spurious
231 leukocytosis and possible determination of its cause. Furthermore, the identification of a cutoff
232 value to differentiate samples with elevated HB could be considered a relevant finding in our
233 study and could be useful in clinical practice. Although we did not evaluate other causes of
234 increased Δ WBCB-WBCP in our study, a value $>1.59 \times 10^9$ WBC/L was strongly related to a
235 marked concentration of circulating HBs in cats.

236 The WBCP appeared to be less affected by the marked HB presence in those cats in our
237 study that had a higher percentage of HBs (Baso-HBs group). To evaluate the accuracy of both
238 automated WBC counts in this population, the automated results were compared to a manual
239 estimation of WBC concentration. This estimation is commonly used in veterinary medicine,
240 particularly in critical care settings. However, its use is scarcely supported by the literature and
241 has inherent limitations that should be taken into account when evaluating the results of our

242 study. Manual WBC estimation was not significantly different from the WBCP, although it was
243 significantly decreased when compared to the WBCB. Nevertheless, these results do not allow us
244 to exclude HB interference on the Perox channel. A previous paper has supported the statement
245 that a manual differential count shows good agreement with the automated lymphocyte and
246 neutrophil differential counts obtained by the analyzer used in our study with the peroxidase
247 reaction.⁴ In our population, however, based on the Perox differential count, significantly
248 increased lymphocyte percentage and absolute number were detected, when compared to the
249 manual differential count results. In our opinion, the spurious increase in the automated
250 lymphocyte count could have been the result of HB interference on the Perox reaction or cell
251 type detection, leading to a misleading classification between lymphocytes and the HBs that
252 were likely identified as small lymphoid cells. For this reason, it is reasonable to hypothesize that
253 the total WBCP could also be overestimated in cats with an elevated percentage of circulating
254 HBs. In these animals, the Perox scattergrams presented dots located in the bottom left-hand
255 corner probably preventing a correct differential leukocyte count, and they were flagged by the
256 PX-NV in 89% of cases, supporting the hypothesis in our study. Similar findings using the laser
257 flow cytometry automated analyzer which preceded the one used in our study have been
258 reported.²⁵ As mentioned above for the WBCB, the visual inspection of the Perox cytogram
259 routinely carried out by the operator could also allow the identification of potential HB-related
260 artifacts on the automated differential WBCP.

261 We also detected an association between the presence of HBs and the artifactual changes
262 in RBC indices. The analyzer that we used has 2 methods for hemoglobin measurement: a
263 cyanide-free colorimetric method (total hemoglobin or mHb) and a flow cytometry cell-based
264 hemoglobin assay (reported as CHCM).³ In contrast to CHCM, the MCHC is a calculated

265 variable that reflects the average concentration of hemoglobin within the erythrocytes, based on
266 the colorimetric measurement of mHb, using the following formula: $MCHC = mHb/[RBC \times$
267 $MCV] \times 1000$. For each sample, MCHC and CHCM are compared automatically, and the system
268 flag CHCMCE is generated if $\Delta MCHC-CHCM$ values are >19 g/L.³ A significant increase in
269 MCHC but not in CHCM (leading to a significant increase in $\Delta MCHC-CHCM$) was observed in
270 the Baso-HBs group. Furthermore, in all of these cats, the RBC scattergrams showed a
271 population of erythrocytes shifted toward hyperchromic RBCs, which could correspond to RBCs
272 with HBs, as suggested previously.²⁴ Thus, the high optical density of numerous RBCs
273 containing HBs in Baso-HBs cats could primarily explain these findings and, particularly, the
274 increased value of the MCHC. It is also known that some other conditions that can cause
275 turbidity in the sample, such as lipemia, icterus, extreme leukocytosis, and in vivo or in vitro
276 hemolysis can result in the overestimation of mHb and MCHC, without a significant effect on
277 CHCM.^{3,23} None of the cats included in our study had macroscopic lipemic or icteric samples,
278 and extreme leukocytosis was excluded by means of manual WBC estimation.

279 In our opinion, the role of hemolysis was of little relevance in the Baso-HBs cats
280 compared to the HBs cats, given the absence of statistical differences in the erythrocyte mass and
281 MCV values, in the morphologic signs of hemolysis other than HBs on blood smear examination
282 (ghost cells and erythrocyte fragments), and in serum total bilirubin concentration (data not
283 shown). However, the possibility that Baso-HBs cats suffered more marked hemolysis could not
284 be completely excluded because Hct and RBCs were slightly decreased in the Baso-HBs group.
285 The absence of statistical significance could be related to the limited number of patients included
286 in the Baso-HBs group and should be considered when interpreting our results.

287 Our results indicate that a value of $\Delta MCHC-CHCM >36$ g/L seemed to be highly

288 predictive of a marked presence of RBCs with HBs, potentially leading to artifactual results in
289 RBC indices. Although CHCM is considered more accurate than mHb or MCHC, HBs might
290 interfere with the light-scattering properties of the erythrocytes affected. This could also cause a
291 mild alteration in the CHCM value, as has been reported.^{23,25} Our results, however, do not
292 support this latter data, given that the CHCM was within the reference interval in all of the cats
293 included, and was not statistically different between the Baso-HBs and HBs cats.

294 The serial CBC results of the Baso-HBs cats evaluated at different times revealed that
295 HBs could persist in peripheral circulation for up to 14 d, with prolonged presence of
296 interference and artifacts in automated CBC and cytogram results. The HB percentage decreased
297 significantly during follow-up evaluation, leading to a reduction in interference on the Baso
298 cytogram and a progressive decrease in WBCB and Δ WBCB-WBCP within 8–14 d from the first
299 clinicopathologic evaluation.

300 Some limitations should be considered when interpreting our results. Given its
301 retrospective design, the medical records of the patients included were not complete in all cases;
302 in particular, inaccurate identification of the underlying disease could have misidentified
303 interference with the automated CBC, in addition to that related to HB formation. Moreover, the
304 sample size was small given the rarity of the artifact herein reported, potentially lowering the
305 power of the statistical analysis carried out.

306 A high percentage of circulating HBs (>36% of RBCs) can lead to significant
307 abnormalities in automated feline CBC results. Even though blood smear examination remains
308 essential in routine veterinary practice to confirm the automated CBC results and identify the
309 presence of HBs, careful evaluation of the cytograms and flags obtained by a flow cytometry–
310 based hematology analyzer could raise suspicion of HB-related interference in cats.

311
312
313
314
315
316
317
318
319
320
321
322
323
324
325
326
327
328
329
330
331
332
333

Acknowledgements

The preliminary results were presented as a poster presentation at the 29th International Society for Laboratory Hematology annual meeting, Milan, 2016.

Declaration of conflicting interests

The authors declared no potential conflict of interest with respect to the research, authorship, and/or publication of this article.

Funding

The authors received no financial support for the research, authorship, and/or publication of this article.

References

1. Allison RW, Meinkoth JH. Hematology without the numbers: in-clinic blood film evaluation. *Vet Clin North Am Small Anim Pract* 2007;37:245–266.
2. Andress JL, et al. The effects of consecutive day propofol anesthesia on feline red blood cells. *Vet Surg* 1995;24:277–282.
3. Bauer N, Moritz A. Evaluation of three methods for measurement of hemoglobin and calculated hemoglobin parameters with the ADVIA 2120 and ADVIA 120 in dogs, cats, and horses. *Vet Clin Pathol* 2008;37:173–179.
4. Bauer N, et al. Evaluation of the automated hematology analyzer Sysmex XT-2000iV™ compared to the ADVIA® 2120 for its use in dogs, cats, and horses. Part II: Accuracy of leukocyte differential and reticulocyte count, impact of anticoagulant and sample aging. *J Vet Diagn Invest* 2012;24:74–89.
5. Christopher MM. Relation of endogenous Heinz bodies to disease and anemia in cats: 120 cases (1978–1987). *J Am Vet Med Assoc* 1989;194:1089–1095.

- 334 6. Christopher MM, et al. Heinz body formation associated with ketoacidosis in diabetic cats. J
335 Vet Intern Med 1995;9:24–31.
- 336 7. Christopher MM, et al. Contribution of propylene glycol-induced Heinz body formation to
337 anemia in cats. J Am Vet Med Assoc 1989;194:1045–1056.
- 338 8. Christopher MM, et al. Erythrocyte pathology and mechanisms of Heinz body-mediated
339 hemolysis in cats. Vet Pathol 1990;27:299–310.
- 340 9. Español I, et al. Heinz bodies interfere with automated reticulocyte counts. Haematologica
341 1999;84:373–374.
- 342 10. Harvey JW. Evaluation of erythrocytes. In: Harvey JW, ed. Veterinary Hematology: A
343 Diagnostic Guide and Color Atlas. 1st ed. St. Louis, MO: Elsevier Saunders, 2012:49–
344 121.
- 345 11. Harvey JW. The feline blood film: 1. Techniques and erythrocyte morphology. J Feline Med
346 Surg 2017;19:529–540.
- 347 12. Hinchliffe RF. Errors in automated reticulocyte counts due to Heinz bodies. J Clin Pathol
348 1993;46:878–879.
- 349 13. Jain N. Demonstration of Heinz bodies in erythrocytes of the cat. Vet Clin Pathol 1973;2:13–
350 23.
- 351 14. Lilliehöök I, Tvedten H. Validation of the Sysmex XT-2000iV hematology system for dogs,
352 cats, and horses. I. Erythrocytes, platelets, and total leukocyte counts. Vet Clin Pathol
353 2009;38:163–174.
- 354 15. Lilliehöök I, Tvedten H. Validation of the Sysmex XT-2000iV hematology system for dogs,
355 cats, and horses. II. Differential leukocyte counts. Vet Clin Pathol 2009;38:175–182.

- 356 16. Matthews NS, et al. Repetitive propofol administration in dogs and cats. *J Am Anim Hosp*
357 *Assoc* 2004;40:255–260.
- 358 17. Moritz A, Becker M. Automated hematology systems. In: Weiss DJ, Wardrop KJ, eds.
359 *Schalm's Veterinary Hematology*. 6th ed. Ames, IA: Blackwell, 2010:1054–1066.
- 360 18. Moritz A, et al. Canine and feline hematology reference values for the ADVIA 120
361 hematology system. *Vet Clin Pathol* 2004;33:32–38.
- 362 19. Rahilly LJ, Mandell DC. Methemoglobinemia. In: Silverstein DC, Hopper K, eds. *Small*
363 *Animal Critical Care Medicine*. 2nd ed. St. Louis, MO: Elsevier Saunders, 2015:580–
364 585.
- 365 20. Reinhart WH, et al. Quantitative relationship between Heinz body formation and red blood
366 cell deformability. *Blood* 1986;68:1376–1383.
- 367 21. Robertson JE, et al. Heinz body formation in cats fed baby food containing onion powder. *J*
368 *Am Vet Med Assoc* 1998;212:1260–1266.
- 369 22. Stirn M, et al. Rate of manual leukocyte differentials in dog, cat and horse blood samples
370 using ADVIA 120 cytograms. *BMC Vet Res* 2014;10:1–8.
- 371 23. Stockham SL, Scott MA. Erythrocytes. In: *Fundamentals of Veterinary Clinical Pathology*.
372 2nd ed. Ames, IA: Blackwell, 2008:107–221.
- 373 24. Tvedten HW. Laboratory and clinical diagnosis of anemia. In: Weiss DJ, Wardrop KJ, eds.
374 *Schalm's Veterinary Hematology*. 6th ed. Ames, IA: Blackwell, 2010:152–161.
- 375 25. Tvedten HW, Holan K. What is your diagnosis? *Vet Clin Pathol* 1996;25:148–154.
- 376 26. Tvedten HW, Moritz A. Reticulocyte and Heinz body staining and enumeration. In: Weiss
377 DJ, Wardrop KJ, eds. *Schalm's Veterinary Hematology*. 6th ed. Ames, IA: Blackwell,
378 2010:1067–1073.

- 379 27. Weiss DJ, et al. 1992. Increased mechanical fragility and intravascular lysis of erythrocytes
380 in cats fed a propylene glycol-containing diet. *Comp Haematol Int* 1992;2:157–161.
- 381 28. Werner LL, et al. Spurious leukocytosis and abnormal WBC histograms associated with
382 Heinz bodies. *Vet Clin Pathol* 1997;26:20.
- 383 29. Zandecki M, et al. Spurious counts and spurious results on hematology analysers: a review.
384 Part II: white blood cells, red blood cells, haemoglobin, red cell indices and reticulocytes.
385 *Int J Lab Hematol* 2007;29:4–20.
- 386

387 **Table 1.** Complete blood count and blood smear evaluation results regarding erythrocytes and
 388 platelets in Baso-HBs and HBs cats.

Variable	Unit	RI	HBs group (n = 23)	Baso-HBs group (n = 9)	p value
HBs	%		8.4 (1.0–52.0)	70.4 ± 17.0	<0.001
	/1,000 RBCs		84 (10–520)	704 ± 170	<0.001
HBs +/n cats	%		14/23 (61%)	0/9	<0.001
HBs ++/n cats	%		7/23 (30%)	1/9 (11%)	<0.001
HBs +++/n cats	%		2/23 (9%)	8/9 (89%)	<0.001
Hct	L/L	0.32–0.48	0.30 (0.10–0.38)	0.23 ± 0.60	NS
RBC	×10 ¹² /L	7.00–11.00	6.08 ± 2.20	5.38 ± 1.70	NS
mHb	g/L	100–162	94 ± 33	105 ± 44	NS
cHb	g/L	101–164	90 ± 31	79 ± 23	NS
HDW	g/L	19.4–24.8	24.7 (17.3–50.0)	47.2 ± 7.6	<0.001
MCH	pg	12.3–16.2	15.7 ± 1.9	19.3 ± 3.4	0.005
MCHC	g/L	316–357	347 ± 19	444 ± 72	<0.001
CHCM	g/L	316–356	334 (285–354)	343 ± 17	NS
ΔMCHC-CHCM	g/L	<19	13 ± 18	101 ± 66	<0.001
CHCMCE/n cats	%		7/23 (30%)	9/9 (100%)	<0.001
MCV	fL	36.0–55.0	45.1 ± 5.6	43.6 ± 5.6	NS
RDW	%	13.0–17.0	15.9 (13.1–22.6)	22.7 ± 5.9	0.003
RP	%	0–0.86	0.45 (0.13–2.33)	1.47 ± 1.23	0.02
RC	×10 ⁹ /L	0–86	26 (8–175)	65 ± 39	0.04
NRBC	/100 WBCs		0 (0–3)	0 (0–7)	NS
PLT	×10 ⁹ /L	150–500	325 ± 162	422 ± 160	NS
MPV	fL	8.0–26.0	16.1 ± 4.7	14.2 ± 2.2	NS
PDW	%	55.0–75.0	62.8 ± 5.6	68.2 ± 5.7	0.008
PCT	%	0.03–0.60	0.48 ± 0.24	0.49 (0.43–0.95)	NS

389 Data are reported as median and range (minimum–maximum) or as mean ± standard deviation,
 390 based on their distribution. CHCM = corpuscular hemoglobin concentration mean; CHCMCE =
 391 comparison error MCHC-CHCM; cHb = calculated hemoglobin; HBs = Heinz bodies; Hct =
 392 hematocrit value; HDW = hemoglobin concentration distribution width; MCH = mean
 393 corpuscular hemoglobin; MCHC = mean corpuscular hemoglobin concentration; MCV = mean
 394 corpuscular volume; mHb = measured hemoglobin; MPV = mean platelet volume; NRBC =
 395 nucleated red blood cell; NS = not significant; PCT = plateletcrit; PDW = platelet volume
 396 distribution width; PLT = platelet count; RBC = red blood cell; RC = reticulocytes number;

397 RDW = red cell volume distribution width; RI = reference interval; RP = reticulocytes
398 percentage; Δ MCHC-CHCM = difference between MCHC and CHCM.

399 **Table 2.** Automated complete blood count and manual count results regarding the WBC in Baso-
 400 HBs and HBs cats.

Variable	Unit	RI	HBs group (n = 23)	Baso-HBs group (n = 9)	p value
WBCB	×10 ⁹ /L	3.9–12.8	11.8 (3.4–56.2)	51.0 (10.6–549.2)	<0.001
WBCP	×10 ⁹ /L	3.7–12.9	12.5 (3.7–54.6)	17.8 (8.2–56.2)	NS
ΔWBCB-WBCP	×10 ⁹ /L	0–0.6	0 ± 1.5	26.7 (2.4–493.1)	<0.001
MC-WBC	×10 ⁹ /L			16.0 ± 10.5	
MN	%	20.8–67.5	18.1 (5.2–89.8)	10.2 ± 10.0	0.02
PMN	%	32.1–79.1	81.7 (10.0–94.6)	89.6 ± 10.0	0.02
Neutrophils	%	30.1–85.4	83.8 (8.5–93.3)	37.9 ± 27.2	0.002
MC neutrophils	%	30.1–85.4	82.0 (25.0–95.0)	86.0 (49.0–92.0)	NS
Lymphocytes	%	6.8–58.2	11.6 (2.7–62.6)	60.2 ± 29.2	<0.001
MC lymphocytes	%	6.8–58.2	13.0 (2.0–74.0)	10.0 (5.0–38.0)	NS
Monocytes	%	0–4.1	2.3 (0.6–44.4)	0.4 (0–4.0)	0.003
MC monocytes	%	0–4.1	2.0 (0–11.0)	2.9 ± 3.4	NS
Eosinophils	%	0.1–13.4	3.5 (0.4–10.8)	0.3 (0–3.8)	<0.001
MC eosinophils	%	6.8–58.2	2.0 (0–22.0)	1.0 (0–4.0)	NS
Basophils	%	0–0.30	0.10 (0–0.40)	0 (0–0.20)	NS
LUC	%	0–0.10	0.10 (0–1.10)	0 (0–0.10)	0.005
Neutrophils	×10 ⁹ /L	1.6–10.0	8.7 (0.8–52.5)	16.9 (7.8–54.3)	NS
MC neutrophils	×10 ⁹ /L	1.6–10.0	8.1 (2.6–50.0)	13.4 ± 9.3	NS
Lymphocytes	×10 ⁹ /L	0.9–5.6	1.4 (0.5–7.6)	27.2 (1.7–493.9)	<0.001
MC lymphocytes	×10 ⁹ /L	0.9–5.6	1.5 (0.3–8.8)	1.5 ± 0.7	NS
Monocytes	×10 ⁹ /L	0–0.65	0.28 (0.07–4.18)	0.26 ± 0.17	NS
MC monocytes	×10 ⁹ /L	0–0.65	0.27 (0–2.81)	0.48 ± 0.56	NS
Eosinophils	×10 ⁹ /L	0.06–1.47	0.43 (0.03–1.46)	0.29 (0–1.44)	NS
MC eosinophils	×10 ⁹ /L	0.06–1.47	0.31 (0–2.98)	0.14 ± 0.16	NS
Basophils	×10 ⁹ /L	0–0.06	0.01 (0–0.23)	0.03 ± 0.03	NS
LUC	×10 ⁹ /L	0–0.01	0.01 (0–0.13)	0.01 (0–0.03)	NS
B-NV/n cats	%		3/23 (13%)	9/9 (100%)	<0.001
PX-NV/n cats	%		3/23 (13%)	8/9 (89%)	<0.001

401 Data are reported as median and range (minimum–maximum) or as mean ± standard deviation,
 402 based on their distribution. B-NV = Baso no valley; ΔWBCB-WBCP = difference between
 403 WBCB and WBCP; HBs = Heinz bodies; LUC = large unstained cell; MC = manual count; MC-
 404 WBC = WBC manual total estimation; MN = mononuclear cell; NS = not significant; PMN =
 405 polymorphonuclear cell; PX-NV = perox no valley; RI = reference interval; WBC = white blood
 406 cell; WBCB = baso total white blood cell count; WBCP = perox total white blood cell count.

407 **Table 3.** Results of the ROC curve analyses evaluating HB%, Δ MCHC-CHCM, and Δ WBCB-
408 WBCP values for the differentiation between cats with or without HB-related artifacts in their
409 Baso cytograms. Cutoff values with maximal sensitivity and specificity for the presence of the
410 artifact are provided.

Variable	Unit	AUC	Cutoff	Sensitivity (%)	Specificity (%)	<i>p</i> value
HBs	%	0.99	>36	100	91	<0.001
Δ MCHC-CHCM	g/L	0.98	>36	100	96	<0.001
Δ WBCB-WBCP	$\times 10^9/L$	0.99	>1.59	100	96	<0.001

411 AUC = area under the curve; Δ MCHC-CHCM = difference between MCHC and CHCM;

412 Δ WBCB-WBCP = difference between WBCB and WBCP; HBs = Heinz bodies; WBC = white

413 blood cell.

414 **Table 4.** Selected hematology variable results obtained from cats belonging to the Baso-HBs group at different times, following
 415 inclusion in the study.

Variable	Unit	T 0 (<i>n</i> = 4)	T1 (1–3 d; <i>n</i> = 4)	T2 (4–7 d; <i>n</i> = 4)	T3 (8–14 d; <i>n</i> = 4)	<i>p</i> value
HBs	%	85.0 (57.4–95.0)	57.9 (34.8–86.0)§	35.6 (25.8–69.8)†	9.4 (2.6–16.6)*	<0.001
Hct	L/L	0.26 (0.15–0.33)	0.22 (0.15–0.23)	0.22 (0.13–0.28)	0.25 (0.13–0.38)	NS
RBC	×10 ¹² /L	7.16 (3.15–7.26)	5.01 (3.15–6.69)§	4.81 (2.61–5.75)§	5.81 (2.68–6.85)	0.03
mHb	g/L	134 (61–184)	100 (60–109)§	82 (49–104)§	80 (48–122)§	0.01
cHb	g/L	90 (52–119)	77 (51–82)	70 (42–94)	73 (42–120)	NS
MCHC	g/L	511 (412–558)	454 (410–488)	378 (399–364)†	325 (318–359)*	<0.001
CHCM	g/L	354 (330–370)	349 (326–387)	334 (314–339)†	313 (290–320)*	<0.001
ΔMCHC-CHCM	g/L	148 (64–223)	90 (60–155)	51 (35–61)†	23 (6–39)*	<0.001
MCV	fL	42.9 (32.5–47.2)	44.7 (31.6–48.7)	47.4 (38.8–53.4)‡	49.1 (34.6–56.2)‡	<0.001
RP	%	1.05 (0.21–1.46)	0.54 (0.31–1.45)	2.30 (2.04–4.67)‡	2.75 (1.74–4.81)‡	<0.001
RC	×10 ⁹ /L	51 (15–94)	28 (17–46)	124 (92–206)‡	143 (79–243)‡	<0.001
WBCB	×10 ⁹ /L	260.6 (41.6–549.2)	171.0 (32.2–278.5)	50.6 (15.7–142.8)	9.7 (8.6–104.0)	NS
WBCP	×10 ⁹ /L	18.9 (9.8–56.2)	8.1 (7.3–21.2)	12.5 (7.8–72.7)	9.0 (6.7–111.4)	NS
ΔWBCB-WBCP	×10 ⁹ /L	241.7 (31.8–493.1)	156.4 (24.0–271.2)	32.0 (7.9–127.3)†	0.6 (–7.4 to 2.0)*	<0.001


416 CHCM = corpuscular hemoglobin concentration mean; cHb = calculated hemoglobin; ΔMCHC-CHCM = difference between MCHC
 417 and CHCM; ΔWBCB-WBCP = difference between WBCB and WBCP; HBs = Heinz bodies; Hct = hematocrit value; MCHC = mean
 418 corpuscular hemoglobin concentration; MCV = mean corpuscular volume; mHb = measured hemoglobin; NS = not significant; RBC =
 419 red blood cell; RC = reticulocytes number; RP = reticulocytes percentage; WBCB = baso total white blood cell count; WBCP = perox
 420 total WBC count.

421 * Significantly different from T0, T1, T2.

422 † Significantly different from T0, T1, T3.

423 ‡ Significantly different from T0 and T1.

424 § Significantly different from T0.



425 **Figure 1. A.** Normal Basophil/lobularity (Baso) cytogram of a cat included in the Heinz bodies
426 (HBs) group. **B.** Baso cytogram of a cat included in the Baso-HBs group. For both graphs, the
427 nuclear lobularity/density is plotted on the x-axis; the cell size is plotted on the y-axis. In B, a
428 large cluster of small particles is visible at the bottom threshold and prevents a clear
429 separation between the mononuclear (MN) and polymorphonuclear (PMN) cell populations
430 that is flagged by Baso no valley (B-NV). Baso+Lyse-resistant = basophils and lyse-resistant
431 cells.

432 **Figure 2. A.** Normal peroxidase (Perox) cytogram of a cat included in the Heinz bodies (HBs)
433 group. **B.** Perox cytogram from a cat of Baso-HBs group. For both graphs, the peroxidase
434 content of the cells is plotted on the x-axis, and the cell size is plotted on the y-axis. In B, a
435 cluster of dots at the bottom left-hand corner prevents the correct separation between cell
436 populations flagged by Perox no valley (PX-NV). Lym = lymphocytes; LUC = large
437 unstained cell; Mon = monocytes; Neu = neutrophils; Agg = platelet aggregates.

438 **Figure 3.** Peripheral blood smear of a cat showing numerous Heinz bodies projecting from the
439 erythrocyte membrane. Brilliant cresyl blue, 600 \times . Bar = 10 μ m.

440 **Figure 4.** Dot plots of mean corpuscular hemoglobin concentration (MCHC g/L; white dots) and
441 corpuscular hemoglobin concentration mean (CHCM g/L; black dots) values in Heinz bodies
442 (HBs) cats ($n = 23$) and in Baso-HBs cats ($n = 9$). The horizontal bars represent the median
443 (central bar) and the minimum and maximum values (peripheral bands); the gray area depicts
444 the reference interval. Significant differences among the variables are indicated.

445 **Figure 5.** Dot plots of Basophil/lobularity white blood cell count (WBCB $\times 10^9/L$; white dots)
446 and peroxidase white blood cell count (WBCP $\times 10^9/L$; black dots) values in Heinz bodies
447 (HBs) cats ($n = 23$) and in Baso-HBs cats ($n = 9$). The horizontal bars represent median

448 (central bar) and minimum and maximum values (peripheral bands); the gray area depicts the
449 reference interval. Significant differences among the variables are indicated.

450 **Figure 6. A.** Normal red blood cell (RBC) cytogram of a cat included in the Heinz bodies (HBs)
451 group. **B.** RBC cytogram from a cat in the Baso-HBs group. For both graphs, the RBC
452 hemoglobin concentration is plotted on the x-axis, and the RBC volume is plotted on the y-
453 axis. In B, the cytogram shows an extension of the RBC cluster to the lower right, indicating a
454 hyperchromic RBC population.