

SARS-CoV-2 inactivation by supercritical carbon dioxide coupled with hydrogen peroxide

Riccardo Zulli^{a,*}, Claudia Del Vecchio^b, Pietro Andrigo^a, Valeria Conciatori^b, Fabio Santi^a, Alessandro Zambon^c, Enrico Lavezzo^b, Sara Spilimbergo^{a,*}

^a Department of Industrial Engineering, University of Padova, Padova, Italy


^b Department of Molecular Medicine, University of Padova, Padova, Italy

^c Department of Civil, Chemical, Environmental, and Materials Engineering, University of Bologna, Bologna, Italy

HIGHLIGHTS

- An experimental setup for supercritical CO₂ without pump was validated.
- SARS-CoV-2 sterilisation was achieved with a small amount of H₂O₂.
- Synergism between supercritical CO₂ and H₂O₂ was verified.

GRAPHICAL ABSTRACT



ARTICLE INFO

Keywords:

SARS-CoV-2
Carbon dioxide
Supercritical fluids
Hydrogen peroxide
COVID-19

ABSTRACT

The recent COVID-19 pandemic has underscored the need for innovative decontamination techniques capable of treating sensitive materials potentially contaminated. Combining supercritical carbon dioxide (scCO₂) with sterilant agents has shown promise in this regard. This study aimed at testing scCO₂ as a virus inactivation method for biomedical materials contaminated with the Severe Acute Respiratory Syndrome Corona Virus 2 (SARS-CoV-2). The virus was inoculated on a stainless-steel carrier and treated at 45 °C and 8 MPa. No inactivation was detected when only scCO₂ was used, even after long treatment times (60 min). The addition of 50 ppm of H₂O₂ to the process allowed the inactivation of more than 5 Log PFU (Plaque Forming Unit) of the virus by only pressurising and depressurising the vessel, while a 20-min process is needed by only using H₂O₂. Overall, the study demonstrates a synergistic effect when H₂O₂ is added to the scCO₂ process for the inactivation of SARS-CoV-2.

* Corresponding authors.

E-mail addresses: riccardo.zulli@unipd.it (R. Zulli), sara.spilimbergo@unipd.it (S. Spilimbergo).

<https://doi.org/10.1016/j.supflu.2024.106272>

Received 28 January 2024; Received in revised form 3 April 2024; Accepted 8 April 2024

Available online 9 April 2024

0896-8446/© 2024 The Authors. Published by Elsevier B.V. This is an open access article under the CC BY license (<http://creativecommons.org/licenses/by/4.0/>).

1. Introduction

The recent pandemic caused by the diffusion of the Severe Acute Respiratory Syndrome Corona Virus 2 (SARS-CoV-2) has resulted in 766.9 million confirmed cases and 6.9 million deaths globally [1]. Studies confirmed that infectious virions remain viable on various kinds of surfaces, especially plastic and steel, for several hours [2], representing a significant factor in virus transmission and human infection. This pandemic has also raised attention to the global production capacity of biomedical instruments, devices and personal protective equipment (PPE), especially face masks, which fell significantly short of meeting the demand of that particular period of time [3]. To address this challenge, the development and optimisation of decontamination technologies to facilitate the quick and safe sanitisation and reuse of biomedical instruments and PPE could not only prevent future unpreparedness in the healthcare industry but also mitigate the significant waste production and environmental impact associated with a similar event [4].

Standard sterilisation methods in the biomedical industry usually include radiation, steam, and disinfectant gases. Radiation-based decontamination techniques, such as gamma, electron beam (E-beam), X-ray, and UV-C, have been demonstrated effective against a large number of human pathogens, including corona viruses [5]. However, the detrimental effects and structural changes observed on different products (i.e. polymers and tissue allografts) treated with radiation-based sterilisation technology [6], together with the limited penetration capacity towards dense material or less-exposed sections of the products [7,8], represent a limitation for their application. On the other hand, the use of steam and disinfectant gases, namely hydrogen peroxide and ethylene oxide, may lead to the alteration in the performance of PPE, such as compromised filtration capacity of face respirators [9], or deposition of residues on the sterilised product. Moreover, a strong effort has been addressed in the last years by the European Medicines Agency (EMA) and the US Food and Drug Administration (FDA) and the Environmental Protection Agency (EPA) to reduce the worker and patient exposure to ethylene oxide, reported to be associated with various cancers, such as lymphoma and leukaemia [10].

In the context of the COVID-19 pandemic, different techniques have been explored to inactivate the virus and some indicator spores on protective masks and other matrices. Biffi et al. [11], for instance, achieved a 3-Log reduction of SARS-CoV-2 through the application of 280 nm UV LED irradiation in a plaque assay test. Hydrogen Peroxide Vapour (HPV) process has been tested for the sterilisation of masks obtaining a total inactivation of *Geobacillus stearothermophilus* [12] and some bacteriophages [9]; however, the cost for the industrial plant was estimated at around 6.9 million USD [13], thus very expensive for its common use on large scale.

Supercritical carbon dioxide (scCO₂) is a promising sterilisation agent that has been extensively studied in recent years. scCO₂ exhibits unique properties, including high diffusivity, low viscosity, and adjustable density, facilitating its penetration and dissolution into diverse materials. Moreover, scCO₂ is non-toxic, non-flammable, and it is considered environmentally friendly, making it a desirable alternative to conventional solvents, particularly in sensitive material processing [14].

The bactericidal effect of scCO₂ was first reported in the 1950s [15], and since then, many research papers have focused on the inactivation of bacteria, yeasts, and moulds using supercritical carbon dioxide for food and biomedical applications, demonstrating its effectiveness for a wide range of matrices and process conditions. In this case, as clearly shown by several authors [14,16,17], the interaction between scCO₂ and the cell membranes, together with the alteration of internal metabolisms, seem to be the main cause of cell death. However, scCO₂ alone is not able to inactivate bacterial spores and viruses because of their higher resistance to thermal and chemical penetration [14] and the use of high temperatures [18] or chemical additives is needed. The coupling effect of CO₂ with different sterilants, such as peracetic acid [19] and

hydrogen peroxide [20], has been investigated obtaining promising results in the sterilisation of spores and viruses. Indeed, according to literature [19,20], the combination of CO₂ and chemical additives can strongly affect the permeability of the cells (outer envelope damage) and alter the internal metabolism by pH reduction, inactivation of proteins and enzymes and introduction of free radicals.

Regarding the sterilisation with scCO₂, the process has been recently applied on different spore strains (e.g., *Bacillus atrophaeus*, *Geobacillus stearothermophilus*, *Bacillus pumilus*) inoculated onto respiratory protective equipment [3,23]. Inactivation to undetectable levels has been achieved using additives, especially hydrogen peroxide, peracetic acid, and ethanol, at concentrations ranging from 100 to 10,000 ppm. Utilising lower amounts of sterilants preserves the physical and chemical attributes of the tested face masks, including their filtration capacity. However, results on the SARS-CoV-2 strain treated with scCO₂ are scarcely present in the literature particularly due to the difficulties in safely performing these tests with high-pressure plants. To the best of the authors' knowledge, only Bennet et al. [21] have examined scCO₂ treatment on SARS-CoV-2 inoculated on N95 respirators, obtaining a complete inactivation after 90 min at 33–35 °C and 100 bar by using 1.25 % (v/v) of a solution composed of ethanol, peracetic acid and hydrogen peroxide in a 600-mL vessel. However, the efficacy of the vaporised solution alone was not verified making difficult to understand the role of scCO₂ in the inactivation mechanism.

This study aims to investigate the inactivation capacity of scCO₂ in synergism with a 50 ppm of H₂O₂ against SARS-CoV-2. This work permits to verify the inactivation behaviour obtained by previous studies on test microorganisms, such as bacterial spores or surrogate viruses, using, for the first time, a concentration of sterilant that has been shown to not alter the properties of the tested materials. Moreover, the inactivation was evaluated considering individual treatments (scCO₂ and H₂O₂ alone) which were overlooked in prior research, in order to demonstrate their synergistic effect.

2. Material and methods

2.1. Cells and virus preparation and inoculation

All the experiments, virus handling, and recovery procedures were carried out in a biosafety level 3 (BSL-3) facility at the Department of Molecular Medicine, University of Padova, Italy.

The African GreenMonkey Kidney Vero E6 (ATCC CRL-1586) cell lines were purchased from the American Type Culture Collection (ATCC) and maintained in Dulbecco's Modified Eagle Medium (DMEM, Euroclone, Milano, Italy) supplemented with 10 % foetal bovine serum (FBS, Euroclone, Milano, Italy), 2 mM glutamine, 1 mM sodium pyruvate, 100 U/mL penicillin, and 100 µg/mL streptomycin sulphate (P/S, both from Life Technologies, Carlsbad, CA, USA). SARS-CoV-2 (Wuhan Hu-1 strain) was obtained from Istituto Superiore della Sanità (ISS) in Rome, Italy. The virus was propagated and titrated in Vero E6 cells following the method described in [24]. Briefly, each viral suspension was prepared and amplified on a large scale in monolayer cell cultures. After infection and multiplication of the virus, cell debris was removed by double centrifugation at low speed (699 × g for 10 min), and the supernatant containing the virus was collected to determine the viral titre (as described in Section 2.3). The supernatant was then aliquoted and stored at – 80 °C.

In order to evaluate the viricidal effect of the method, the virus suspension was tested following the standard test on non-porous surface without mechanical action EN 1677 [25]. Before each test, a volume of 50 µL of the viral suspension (around 10⁷ PFU/mL, Plaque Forming Units per millilitre) was directly inoculated on sterile stainless-steel discs with a diameter of 35 mm, used as non-porous test carriers. The suspension was allowed to dry for around 30 min under a laminar flow hood. The control carriers were inoculated with the viral suspension (50 µL) but not subjected to any treatment.

2.2. Treatment procedure

The high-pressure tests were performed using a stainless-steel vessel with bolted closure, an internal volume of 330 mL and an internal height and diameter of 11 cm and 6 cm respectively, and able to hold up to 30 MPa. The vessel was equipped with a temperature probe (PT100, National Instruments, Milano, Italy) and a pressure gauge (EN837, Wika, Milano, Italy) to monitor the internal temperature and pressure during the treatments. The temperature probe was positioned near the inoculated disc at the top of the vessel. The vessel presented one outlet tube with a manual micrometric valve (Swagelok, Nordival, Milano, Italy) that was maintained closed until the depressurisation phase. A schematic representation of the setup is reported in Fig. 1.

The vessel was autoclaved and pre-heated to 45 °C before each test. The experimental procedure was performed in a BSL-3 facility, thus a specific setup was implemented and optimised for the study. Usually, scCO₂ processes are performed within a high-pressure setup, which includes several components (e.g. CO₂ tank, tubes, high-pressure pump, chiller). For safety reasons, it was not allowed to include a conventional setup in a BSL-3 facility. The supercritical conditions were achieved by using solid CO₂ followed by a change of state due to temperature increase and its consequent increase of pressure induced by a controlled volume. In particular, the amount of CO₂ inserted (115 g) was optimised to achieve the desired pressure conditions during the process considering thermodynamics calculation and practical arrangements (see Section 3). After the addition of CO₂, the inoculated disc was placed on a stainless-steel stand and the vessel was quickly closed. The reduction in temperature caused by the solid CO₂ was compensated by using a heating plate positioned under the vessel, leading to the achievement of supercritical conditions. Temperature and pressure were maintained at the target values of 45 ± 0.5 °C and 8.0 ± 0.1 MPa, respectively, ensuring the supercritical state of CO₂. The treatment time was defined as the duration from reaching the desired pressure until the initiation of depressurisation. The pressurisation step lasted approximately 2 min, while the depressurisation time was fixed at 3 min by manually opening the micrometric valve.

Mixing during the treatments was ensured by a magnetic stirrer




Fig. 1. Schematic representation of the experimental setup.

positioned at the bottom of the vessel, operating at 400 rpm. High-pressure treatments were conducted with (scCO₂ + H₂O₂) or without (scCO₂) the addition of 15 µL (50 ppm) hydrogen peroxide (H₂O₂) (30 % wt in water, Sigma-Aldrich, Burlington, MA, USA). To evaluate the combined effects of CO₂ and H₂O₂, treatments with only temperature (CTRL-T) and treatments with both temperature and H₂O₂ (H₂O₂) were included.

2.3. Determination of the viral titre

The infectivity titres of the virus for both untreated and treated samples were recovered by eluting the stainless-steel discs with 1 mL of culture medium and scraping them for 1 min. Serial dilutions ranging from 10⁻² to 10⁻⁹ (1:10 dilutions) were prepared.

Briefly, 10⁵ Vero E6 cells/well were seeded in a 24-well plate in 500 µL/well of DMEM 10 % FBS. The day after, when the confluence was roughly 90 %, cells were infected with 250 µL of 10-fold dilutions (10⁻²–10⁻⁹) of the virus of the untreated and treated samples in serum-free DMEM. After an incubation of 1 h at 37 °C, cells were washed with PBS (Phosphate Buffered Saline) (1X, pH 7.4) to remove the cell-free virus and incubated at 37 °C with 500 µL/well DMEM supplemented with 2 % v/v FBS and 0.75 % v/v carboxy-methyl-cellulose.

After the incubation at 37 °C for 72 h, plaque formation was assessed to determine PFU/mL. Cells were washed, fixed with 500 µL of 5 % formaldehyde for 10 min and stained with 0.1 % v/v crystal violet (for 50 mL: 50 mg crystal violet, 10 % methanol and deionized H₂O to volume). The plaques were counted at the optical microscope and the viral titre was calculated using the following equation:

$$X = n / (D * V) \quad (1)$$

where X is the virus titre in PFU/mL, n is the average number of plaques counted, D is the dilution factor and V is the volume of diluted virus/well. A decrease of ≥ 4 log units of viral titre was considered a viricidal effect.

2.4. Statistical analysis

Minitab® software was used to analyse statistical differences between the experimental and control groups. After verifying the independence of the observations, the normality of the data and the homogeneity of variances, ANOVA with post-hoc Tukey HSD analysis was performed, and a p-value of 0.05 was used as the threshold for significance. In order to verify the reproducibility, each treatment was repeated at least two times and the same sample was analysed twice.

3. Results and discussion

3.1. Experimental setup optimisation

In this study, the inactivation efficacy of scCO₂ in combination with hydrogen peroxide was assessed on stainless-steel discs inoculated with SARS-CoV-2. Preliminary trials were conducted to optimise and validate a new experimental setup that avoids the use of CO₂ tanks and pressurisation pumps by acting only on the temperature control. The protocol was optimised to ensure a robust experimental procedure using solid CO₂ and achieving the supercritical conditions in a short time. To compensate for the heat losses and temperature decrease during the CO₂ sublimation, the vessel was preheated and positioned on a hot plate. The temperature was fixed at 45 °C and monitored by a temperature probe located close to the inoculated disc. Different amounts of solid CO₂ were tested (data not shown) and 115 g of CO₂ was found to be the optimal initial amount to achieve a stable pressure of 8.0 MPa after only 2 min from the sealing of the vessel. Notably, the quantity used is higher than the estimation based on calculations by vessel internal volume (330 mL) and carbon dioxide density under the given conditions (250.46 kg/m³, [26]). This discrepancy might be attributed to the loss of gaseous CO₂

through sublimation during the initial protocol phase (weight of solid CO₂, insertion inside the vessel, disc positioning and sealing of the vessel lid). Larger quantities of solid CO₂, which could have allowed the obtainment of higher pressure, were not considered due to constraints related to vessel volume and the need to avoid the direct contact of solid CO₂ with the inoculated disc.

The stirrer velocity was also optimised and fixed at 400 rpm to allow a quick distribution of CO₂ and the additive while ensuring its correct rotation inside the vessel. Fig. 2 reports the evolution of temperature and pressure overtime for the optimised protocol at 45 °C.

After the closure of the vessel, pressure rapidly increases to the desired value (8 MPa) in about 2 min with a slight deceleration near the critical pressure. The pressure set point is maintained during the whole treatment. At the end of the processing time, depressurisation can be achieved in about 3 min by opening the exit valve, corresponding to a depressurisation rate that remains roughly constant at around 2.67 MPa/min. The temperature, which was initially pre-heated at 45 °C, is partially cooled down due to the rapid change of phase of solid CO₂. However, the hot plate rapidly compensates for this decrease in temperature and mitigates heat losses during the treatment, ensuring a stable temperature profile. During depressurisation, the temperature quickly decreases due to the rapid decompression of CO₂.

The validated experimental setup allowed us to test the high-pressure technology inside a BSL-3 facility without the use of bulky and expensive equipment, such as high-pressure pumps, CO₂ tanks and chillers.

3.2. Inactivation tests

Once the operative protocols were optimised, stainless-steel discs were inoculated starting from an initial inoculum with a viral titre of 6.14 ± 0.29 Log PFU and the dried samples were subjected to treatments involving scCO₂ and/or hydrogen peroxide. Table 1 reports the obtained results for both control and experimental runs.

Temperature (45 °C) and supercritical carbon dioxide (45 °C and 8 MPa) alone did not result in a significant reduction of viral titre, even after a 60-min exposure period. These findings are in agreement with the previous literature studies. Indeed, different studies investigated the effect of low-temperature scCO₂ on various strains of viruses, such as vesicular stomatitis virus (VSV) [13]. Similar results were achieved also for the inactivation of bacterial spores, including *Bacillus subtilis* [18] and *Bacillus coagulans* [27], but the effects were always minimal unless high temperatures (> 60 °C) [18], and/or extended treatment times




Fig. 2. Temperature (blue) and pressure (orange) profile of the optimised protocol. The vertical dashed lines represent the start and the endpoint of the treatment, respectively.

Table 1

SARS-CoV-2 viral titre on the control carrier (CTRL) and after treatment by only temperature at 45 °C (CTRL-T), by scCO₂ at 45 °C and 8 MPa (scCO₂), by H₂O₂ at 45 °C (H₂O₂), and by scCO₂ and H₂O₂ at 45 °C and 8 MPa (scCO₂ + H₂O₂). Data are expressed in Log PFU (mean \pm standard deviation). Each test was repeated at least twice. Means with different superscript letters are statistically different (p-value \leq 0.05).

Sample	Time (min)	Viral titre (Log PFU)
CTRL	-	5.13 ± 0.39^a
CTRL-T	30	5.16 ± 0.15^a
	60	4.74 ± 0.09^{ab}
scCO ₂	30	4.82 ± 0.17^{ab}
	60	4.37 ± 0.20^{ab}
H ₂ O ₂ (50 ppm)	10	4.60 ± 0.16^{ab}
	15	4.00 ± 0.03^b
	20	n.d.
scCO ₂ + H ₂ O ₂ (50 ppm)	0	n.d.
	5	n.d.
	10	n.d.

n.d. = non-detected, viral titre < 1 Log PFU.

(days) were adopted [28].

Given its strong oxidising properties and its confirmed spore and virus inactivation capabilities, hydrogen peroxide was selected as an additional agent. Moreover, it is an environmentally safer agent with respect to other sterilants usually adopted, such as peracetic acid or ethylene oxide [29,30]. Firstly, experiments were conducted to assess the efficacy of hydrogen peroxide alone at 45 °C. In the selected experimental setup, a concentration of 50 ppm of hydrogen peroxide at 45 °C needed 20 min to completely inactivate the virus.

Similar tests have been conducted by Goyal et al. [31] on various virus strains, including Transmissible gastroenteritis coronavirus (TGEV) (a SARS-CoV surrogate) inoculated on a stainless steel disc, with a much higher concentration of hydrogen peroxide and longer treatment times. The authors exposed the samples to hydrogen peroxide vapour (HPV) in an environmental chamber with recirculation inactivating to undetectable levels with a minimum sterilant volume of 25 mL (2 mL/min for 1 min followed by 1.5 mL/min for 15 min) in a 4-L vessel and with a total exposure time, including injection and aeration, of approximately 2–3 h.

When scCO₂ treatment was combined with H₂O₂, the pressurisation and depressurisation of the vessel alone were sufficient to completely inactivate the virus on the stainless-steel disc, resulting in a substantial reduction in inactivation time. Similarly, Bennet et al. [21] achieved a complete inactivation of SARS-CoV-2 inoculated on N95 respirators by employing a combination of scCO₂ and sterilant additives. In their study, samples were subjected to treatment at 33–35 °C and 10 MPa for a duration of 90 min. Despite the lower temperature, their approach involved prolonged exposure and a considerably higher sterilant volume compared to this study: 7.5 mL of a solution composed of ethanol, peracetic acid and hydrogen peroxide in a 600-mL vessel (> 10,000 ppm). Furthermore, the exclusive application of vaporised sterilant to evaluate the synergistic effect with scCO₂ was not assessed, making it challenging to ascertain the potential synergy between the agents.

To the best of the author's knowledge, there are no other published studies about SARS-CoV-2 inactivation by scCO₂ methods. Other virus strains have been tested, such as *Encephalomyocarditis virus* (EMCV). Qiu et al. [19] treated EMCV-inoculated porcine acellular matrix samples with scCO₂ and peracetic acid obtaining a complete inactivation after 15 min at 10 MPa and 37–41 °C and 55 ppm of sterilant. The longer inactivation time can be attributed to the slightly lower temperature, the system setup and the strain used. EMCV is indeed a type of non-enveloped virus, usually more resistant with respect to enveloped ones, such as SARS-CoV-2. Thus, the fast inactivation obtained in this study could be imputed to the lipophilic nature of the scCO₂ and its interaction and solubilisation with the virus envelope, causing membrane disruption and, therefore, the virus inactivation [22].

4. Conclusions

This study focused on the inactivation of SARS-CoV-2 virus inoculated on stainless steel discs using supercritical carbon dioxide (scCO₂) coupled with hydrogen peroxide (H₂O₂). The virus was effectively inactivated after only pressurising and depressurising the vessel with 50 ppm of H₂O₂. The application of the scCO₂ or H₂O₂ alone were not effective even after 60 min. These findings hold significant promise for the development of systems for treating different matrices, particularly for biomedical applications such as the treatment of facial masks for reutilisation. Further studies are needed to optimise the process parameters. In particular, the sterilant amount could be further reduced in order to decrease the risks of side effects of the treatment on the treated matrices and increase the range of products that could be treated with this method, such as natural and synthetic materials for tissue engineering. Moreover, different strains of SARS-CoV-2 and/or other viruses, such as non-enveloped ones, could be tested with the experimental setup used in this study.

Funding

This research did not receive any specific grant from funding agencies in the public, commercial, or not-for-profit sectors.

CRediT authorship contribution statement

Riccardo Zulli: Writing – review & editing, Writing – original draft, Visualization, Validation, Methodology, Investigation, Formal analysis, Data curation, Conceptualization. **Claudia Del Vecchio:** Writing – review & editing, Writing – original draft, Visualization, Validation, Supervision, Methodology, Investigation, Data curation, Conceptualization. **Pietro Andriago:** Writing – review & editing, Writing – original draft, Visualization, Methodology, Investigation, Formal analysis, Data curation, Conceptualization. **Valeria Conciatori:** Writing – review & editing, Writing – original draft, Methodology, Investigation, Conceptualization. **Fabio Santi:** Writing – review & editing, Writing – original draft, Methodology, Investigation, Data curation, Conceptualization. **Alessandro Zambon:** Writing – review & editing, Writing – original draft, Supervision, Methodology, Conceptualization. **Enrico Lavezzo:** Writing – review & editing, Supervision, Methodology, Conceptualization. **Sara Spilimbergo:** Writing – review & editing, Writing – original draft, Visualization, Validation, Supervision, Methodology, Investigation, Conceptualization.

Declaration of Competing Interest

The authors declare that they have no known competing financial interests or personal relationships that could have appeared to influence the work reported in this paper.

Data availability

Data will be made available on request.

Acknowledgments

The graphical abstract and Fig. 1 have been created with BioRender.com.

References

- [1] WHO Coronavirus (COVID-19) Dashboard, n.d. (<https://covid19.who.int>) (Accessed 24 May 2024).
- [2] N. van Doremalen, T. Bushmaker, D.H. Morris, M.G. Holbrook, A. Gamble, B. N. Williamson, A. Tamin, J.L. Harcourt, N.J. Thornburg, S.I. Gerber, J.O. Lloyd-Smith, E. de Wit, V.J. Munster, aerosol and surface stability of SARS-CoV-2 as compared with SARS-CoV-1, *N. Engl. J. Med.* 382 (2020) 1564–1567, <https://doi.org/10.1056/NEJMc2004973>.
- [3] V. Santos-Rosales, C. López-Iglesias, A. Sampedro-Viana, C. Alvarez-Lorenzo, S. Ghazanfari, B. Magariños, C.A. García-González, Supercritical CO₂ sterilization: an effective treatment to reprocess FFP3 face masks and to reduce waste during COVID-19 pandemic, *Sci. Total Environ.* 826 (2022) 154089, <https://doi.org/10.1016/j.scitotenv.2022.154089>.
- [4] A. Mahmoudnia, N. Mehrdadi, F. Golbabaee Kootenaee, M. Rahmati Deiranloee, E. Al-e-Ahmad, Increased personal protective equipment consumption during the COVID-19 pandemic: an emerging concern on the urban waste management and strategies to reduce the environmental impact, *J. Hazard. Mater. Adv.* 7 (2022) 100109, <https://doi.org/10.1016/j.hazadv.2022.100109>.
- [5] A. Kumar, A. Raj, A. Gupta, S. Gautam, M. Kumar, H. Bherwani, A. Anshul, Pollution free UV-C radiation to mitigate COVID-19 transmission, *Gondwana Res.* 114 (2023) 78–86, <https://doi.org/10.1016/j.gr.2022.07.010>.
- [6] C.R. Harrell, V. Djonov, C. Fellabaum, V. Volarevic, Risks of using sterilization by gamma radiation: the other side of the coin, *Int. J. Med. Sci.* 15 (2018) 274–279, <https://doi.org/10.7150/ijms.22644>.
- [7] J.C. Rubio-Romero, M. del C. Pardo-Ferreira, J.A. Torrecilla-García, S. Calero-Castro, Disposable masks: disinfection and sterilization for reuse, and non-certified manufacturing, in the face of shortages during the COVID-19 pandemic, *Saf. Sci.* 129 (2020) 104830, <https://doi.org/10.1016/j.ssci.2020.104830>.
- [8] N.P. Tipnis, D.J. Burgess, Sterilization of implantable polymer-based medical devices: a review, *Int. J. Pharm.* 544 (2018) 455–460, <https://doi.org/10.1016/j.ijpharm.2017.12.003>.
- [9] P.A. Kenney, B.K. Chan, K.E. Kortright, M. Cintron, M. Russi, J. Epright, L. Lee, T. J. Balczak, N.L. Havill, R.A. Martinello, Hydrogen peroxide vapor decontamination of N95 respirators for reuse, *Infect. Control Hosp. Epidemiol.* 43 (2022) 45–47, <https://doi.org/10.1017/ice.2021.48>.
- [10] Ethylene Oxide - Cancer-Causing Substances - NCI, 2015. (<https://www.cancer.gov/about-cancer/causes-prevention/risk/substances/ethylene-oxide>) (Accessed 29 February 2024).
- [11] S. Biffi, L. Signorini, L. Cattaneo, L. Della Corna, A. Guercilena, S. D'Alessandro, P. Ferrante, S. Delbue, Determination of the UV inactivation constant under 280 Nm UV LED irradiation for SARS-CoV-2, photochemistry and photobiology, n/a, n.d. (<https://doi.org/10.1111/php.13653>).
- [12] Final Report for the Bioquell Hydrogen Peroxide Vapor (HPV) Decontamination for Reuse of N95 Respirators, Final Report, 2016, p. 46.
- [13] J. Koplou, K. Smith, N. Jouravel, G. Buffleben, A. Sinha, O. Negrete, T. Barnett, R. Karnesky, Supercritical CO₂ sterilization of N95 Masks, Sandia National Lab. (SNL-NM), Albuquerque, NM (United States), 2020. (<https://doi.org/10.2172/1821584>).
- [14] N. Ribeiro, G.C. Soares, V. Santos-Rosales, A. Concheiro, C. Alvarez-Lorenzo, C. A. García-González, A.L. Oliveira, A new era for sterilization based on supercritical CO₂ technology, *J. Biomed. Mater. Res. Part B: Appl. Biomater.* 108 (2020) 399–428, <https://doi.org/10.1002/jbm.b.34398>.
- [15] D. Fraser, Bursting bacteria by release of gas pressure, *Nature* 167 (1951) 33–34, <https://doi.org/10.1038/167033b0>.
- [16] S. Spilimbergo, A. Bertucco, G. Basso, G. Bertoloni, Determination of extracellular and intracellular pH of *Bacillus subtilis* suspension under CO₂ treatment, *Biotechnol. Bioeng.* 92 (2005) 447–451, <https://doi.org/10.1002/bit.20606>.
- [17] L. Garcia-Gonzalez, A.H. Geeraerd, S. Spilimbergo, K. Elst, L. Van Ginneken, J. Debevere, J.F. Van Impe, F. Devlieghere, High pressure carbon dioxide inactivation of microorganisms in foods: the past, the present and the future, *Int. J. Food Microbiol.* 117 (2007) 1–28.
- [18] S. Spilimbergo, A. Bertucco, F.M. Lauro, G. Bertoloni, Inactivation of *Bacillus subtilis* spores by supercritical CO₂ treatment, *Innov. Food Sci. Emerg. Technol.* 4 (2003) 161–165.
- [19] Q.-Q. Qiu, P. Leamy, J. Brittingham, J. Pomerleau, N. Kabaria, J. Connor, Inactivation of bacterial spores and viruses in biological material using supercritical carbon dioxide with sterilant, *J. Biomed. Mater. Res. B Appl. Biomater.* 91 (2009) 572–578, <https://doi.org/10.1002/jbm.b.31431>.
- [20] A. Bernhardt, M. Wehr, B. Paul, T. Hochmuth, M. Schumacher, K. Schütz, M. Gelinsky, Improved sterilization of sensitive biomaterials with supercritical carbon dioxide at low temperature, *PLoS One* 10 (2015) e0129205, <https://doi.org/10.1371/journal.pone.0129205>.
- [21] D. Bennet, A.F. Harris, J. Lacombe, C. Brooks, N. Bionda, A.D. Strickland, T. Eisenhut, F. Zenhausern, Evaluation of supercritical CO₂ sterilization efficacy for sanitizing personal protective equipment from the coronavirus SARS-CoV-2, *Sci. Total Environ.* 780 (2021) 146519, <https://doi.org/10.1016/j.scitotenv.2021.146519>.
- [22] M. Perrut, Sterilization and virus inactivation by supercritical fluids (a review), *J. Supercrit. Fluids* 66 (2012) 359–371.
- [23] A. Cario, G. Aubert, J.-P. Alcaraz, J.-P. Borra, N. Jidenko, M. Barrault, S. Bourroux, V. Mocho, F.X. Ouf, S. Marre, P. Cinquin, C. Aymonier, Supercritical carbon dioxide-based cleaning and sterilization treatments for the reuse of filtering facepiece respirators FFP2 in the context of COVID-19 pandemic, *J. Supercrit. Fluids* 180 (2022) 105428, <https://doi.org/10.1016/j.supflu.2021.105428>.
- [24] A. Calistri, A. Luginani, B. Mognetti, E. Elder, G. Sibille, V. Conciatori, C. Del Vecchio, S. Sainas, D. Boschi, N. Montserrat, A. Mirazimi, M.L. Lolli, G. Gribaudo, C. Parolin, The new generation hDHODH inhibitor MEDS433 hinders the in vitro replication of SARS-CoV-2 and other human coronaviruses, *Microorganisms* 9 (2021) 1731, <https://doi.org/10.3390/microorganisms9081731>.
- [25] E.C. for Standardization—CEN, En 16777: 2019: Chemical Disinfectants and Antiseptics—Quantitative Non-Porous Surface Test Without Mechanical Action for the Evaluation of Virucidal Activity of Chemical Disinfectants Used in the Medical

- Area—Test Method and Requirements (Phase 2/Step 2), European Committee for Standardization Brussels, Belgium, 2019.
- [26] U. Sievers, Die thermodynamischen Eigenschaften von Kohlendioxid, 1984. (<https://doi.org/10.1007/BF02560841>) (Accessed 28 February 2023).
- [27] H. Ishikawa, M. Shimoda, K. Tamaya, A. Yonekura, T. Kawano, Y. Osajima, Inactivation of Bacillus spores by the supercritical carbon dioxide micro-bubble method, *Biosci. Biotechnol. Biochem.* 61 (1997) 1022–1023, <https://doi.org/10.1271/bbb.61.1022>.
- [28] J. Zhang, N. Dalal, C. Gleason, M.A. Matthews, L.N. Waller, K.F. Fox, A. Fox, M. J. Drews, M. LaBerge, Y.H. An, On the mechanisms of deactivation of Bacillus atrophaeus spores using supercritical carbon dioxide, *J. Supercrit. Fluids* 38 (2006) 268–273.
- [29] F. Bounoure, H. Fiquet, P. Arnaud, Comparison of hydrogen peroxide and peracetic acid as isolator sterilization agents in a hospital pharmacy, *Am. J. Health Syst. Pharm.* 63 (2006) 451–455, <https://doi.org/10.2146/ajhp040451>.
- [30] V.F. Scharf, B. Dent, M.E. Jacob, B. Moore, Efficacy of vaporized hydrogen peroxide for repeated sterilization of a single-use single-incision laparoscopic surgery port, *Vet. Surg.* 48 (2019) 059–065, <https://doi.org/10.1111/vsu.12905>.
- [31] S.M. Goyal, Y. Chander, S. Yezli, J.A. Otter, Evaluating the virucidal efficacy of hydrogen peroxide vapour, *J. Hosp. Infect.* 86 (2014) 255–259, <https://doi.org/10.1016/j.jhin.2014.02.003>.